

2012 Quality Assurance Report Atmospheric Mercury Network



National Atmospheric Deposition Program

For information about the National Atmospheric Deposition Program (NADP) contact:

NADP Program Office
Illinois State Water Survey
University of Illinois at Urbana-Champaign
2204 Griffith Drive
Champaign, Illinois 61820-7495

URL: <http://nadp.isws.illinois.edu>

e-mail: nadp@isws.illinois.edu

phone: 217-333-7871

Abbreviations

| | |
|--------------------|--|
| AMNet | Atmospheric Mercury Network |
| GEM | Gaseous Elemental Mercury (expressed in ng/m ³) |
| GOM | Gaseous Oxidized Mercury (expressed in pg/m ³) |
| MDN | Mercury Deposition Network |
| NADP | National Atmospheric Deposition Program |
| PBM _{2.5} | Particulate-Bound Mercury less than 2.5 μm in diameter (expressed in pg/m ³) |
| QAP | Quality Assurance Program |
| SOP | Standard Operating Procedures |

Units and Conversion Factors

| | |
|-------------------|--|
| ° | degrees |
| °C | degrees Celsius |
| cm | centimeters |
| L | liters |
| lpm | liters per minute |
| ng | nanograms (1 ng = 10 ⁻⁹ g) |
| ng/m ³ | nanograms per cubic meter |
| pg | picograms (1 pg = 10 ⁻¹² g) |
| pg/m ³ | picograms per cubic meter |

Table of Contents

| | |
|--|----|
| 1.0 Introduction | 6 |
| 2.0 Site Performance and Systems Surveys | 7 |
| 2.1 AMNet Sites Surveyed in 2012 | |
| 2.2 Instrument Test Results | |
| 2.3 Siting Criteria | |
| 2.4 Instrument Repairs | |
| 3.0 Training | 11 |
| 4.0 Data | 12 |
| Appendix Test Equipment Calibration Documents | 13 |

List of Tables

| | | |
|---------|--|----|
| Table 1 | AMNet Sites Surveyed in 2012 | 8 |
| Table 2 | Serial Numbers for Instruments at Surveyed Sites | 8 |
| Table 3 | Survey Results | 10 |
| Table 4 | Siting Criteria Obstructions and Inlet Heights | 11 |
| Table 5 | Percent Valid Data by Site for 2012 | 12 |

1.0 Introduction

The Atmospheric Mercury Network (AMNet) started in 2009. In 2012 the network consisted of 19 sites across North America. The concentrations of gaseous elemental mercury (GEM), gaseous oxidized mercury (GOM) and particulate bound mercury (PBM_{2.5}) are measured at each site following the AMNet Standard Operating Procedures (SOPs). The AMNet Site Liaison provides remote technical support to site operators in the operation of AMNet equipment, performs site performance and systems surveys, and reviews the data on a monthly basis to identify problems. Data review includes both manual and automated quality control checks. Site operators are notified whenever problems are discovered.

In 2012 twelve sites were surveyed by the AMNet Site Liaison. This report includes a summary of the findings from each of the surveys.

Changes in 2012 include the following:

MD08 (Frostburg, MD) discontinued operations on 01/03/2012

MS12 (Grand Bay, MS) collocated measurements stopped on 11/12/2012

MS12 (Grand Bay, MS) Ron Cole replaced Jake Walker as site operator in October 2012

MS99 (Grand Bay, MS – collocated instruments) Ron Cole replaced Jake Walker as site operator in October 2012

NS01 (Kejimikujik National Park) Rob Keenan replaced John Dalziel as site operator in September 2012

OH02 (Athens, OH) discontinued operations on 02/15/2012

VT99 (Underhill, VT) Mim Pendleton replaced Eric Miller as site operator in March 2012

WV99 (Canaan Valley, WV) discontinued operations on 10/14/2012

Changes to data in 2012 include the following:

NY06 (Bronx) GOM concentrations were adjusted by a factor of 4.167 for the period 01 January through 30 July (at 1820). Flag 4 was changed in the instrument BIOS, but the scale factor had not been changed.

NY95 (Rochester – collocated instrument) GOM concentrations were adjusted by a factor of 4.167 for the period 01 January through 12 July (at 1444). Flag 4 was changed in the instrument BIOS, but the scale factor had not been changed.

OK99 (Stilwell, OK) PBM and GOM concentrations were adjusted by a factor of 4.167 for the period 06 February (at 1320) through 20 September (at 2152). When the instrument was repaired all scale factors were reset in the instrument BIOS to 1.0 by mistake.

VT99 (Underhill, VT) GOM concentrations were adjusted by a factor of 5.21 for the period 01 January through 06 August. Flag 4 was changed in the instrument BIOS, but the scale factor had not been changed.

2.0 Site Performance and Systems Surveys

Sites are surveyed at least once every two years by the AMNet Site Liaison. Normally, the site performance and systems surveys would be performed by an independent entity. This is true for the other four NADP networks. The expertise required to operate and troubleshoot the AMNet instrumentation prohibits an independent third party from providing this service. Field survey reports are completed to document problems that are discovered and their resolution.

Site surveys evaluate both field and laboratory operations (including equipment operation), and siting criteria. Site surveys ensure data comparability within the network, resolve operational problems that may not be apparent in data review, and address training needs at each site.

Additional information regarding site surveys may be found in the document titled *Atmospheric Mercury Network: Site Performance and Systems Survey*. This document is available from the NADP website (<http://nadp.isws.illinois.edu/>).

2.1 AMNet Sites Surveyed in 2012

Site surveys were conducted at twelve AMNet sites in 2012. Station ID's, survey dates and station names are presented in Table 1. Two site IDs are associated with the sites at Grand Bay, MS and Beltsville, MD. These sites operated collocated instruments for at least a portion of the year.

Table 1. AMNet Sites Surveyed in 2012.

| Site ID | Station Name | Survey Date |
|-----------|----------------------------|-------------|
| CA48 | Elkhorn Slough | 1/9/2012 |
| HI00 | Mauna Loa | 1/13/2012 |
| MD08 | Piney Reservoir | 7/3/2012 |
| MD96/MD97 | Beltsville | 7/9/2012 |
| MS12/MS99 | Grand Bay NERR | 3/7/2012 |
| NY06 | Bronx | 7/11/2012 |
| NY20 | Huntington Wildlife Forest | 8/8/2012 |
| NY95 | Rochester | 7/12/2012 |
| OH02 | Athens | 7/2/2012 |
| UT97 | Salt Lake City | 6/12/2012 |
| VT99 | Underhill | 8/6/2012 |
| WV99 | Canaan Valley Institute | 7/5/2012 |

2.2 Instrument Test Results

As part of the site survey, instrument sensitivity (i.e., response factor) and the internal calibration source are verified. Independent, third party calibration certificates for the survey test equipment are included in the appendix to this document.

Table 2 lists the serial numbers for the AMNet instruments at each site. Illegible serial numbers are listed as “n/a” (not available).

Table 2. Serial Numbers for Instruments at Surveyed Sites.

| Site ID | 1102 | 2537 | 1130P | 1130 | 1135 | 2505 |
|---------|---|------|-------|------|------|------|
| CA48 | 86 | 167 | 54 | 54 | n/a | n/a |
| HI00 | 53 | 130 | 18 | 18 | 9 | 51 |
| MD08 | Removed from service, no instrumentation at site. | | | | | |
| MD96 | 25 | 342 | 88 | n/a | n/a | 151 |
| MD97 | 25 | 314 | 82 | 88 | 74 | 151 |
| MS12 | 36 | 254 | 69 | 66 | 53 | 147 |
| MS99 | 36 | 291 | 78 | 67 | n/a | 147 |
| NY06 | n/a | 327 | 84 | n/a | n/a | n/a |
| NY20 | 35 | 321 | 57 | n/a | 46 | n/a |
| NY95 | 46 | 326 | 83 | n/a | n/a | n/a |
| OH02 | 54 | 174 | 47 | 47 | 36 | 81 |
| UT97 | 77 | 364 | 105 | 103 | 88 | 169 |
| VT99 | 22 | 178 | 53 | n/a | 94 | 97 |
| WV99 | 105 | 365 | 57 | n/a | n/a | n/a |

Table 3 lists the results [i.e., pass (p), fail (f)] for each test of the field instruments. Criteria for assigning pass/fail are defined in *Atmospheric Mercury Network: Site Performance and Systems Survey*. Significant deviation from the test criteria are indicated with an uppercase F. Extensive equipment repairs were required at some sites. As a result, there was insufficient time to complete a full survey of the site. Parameters that were not tested are listed as “n/a.”

Table 3. Survey Results.

| Site ID | Survey Date | Air Flow and Leak Tests | | | | Cartridge A and B Recoveries | | | |
|---------|-------------|-------------------------|------------|------------------|----------------|------------------------------|-----------|------------|-------------|
| | | Temps OK | Inlet Flow | 2537 Flow | Leak Check | Response Factor | Low Level | High Level | Ambient Air |
| CA48 | 1/9/2012 | p | n/a | n/a | n/a | p | p | p | p |
| HI00 | 1/13/2012 | p | p | p | p | p | p | p | f |
| MD08 | 7/3/2012 | Removed from service | | | | | | | |
| MD96 | 7/9/2012 | p | p | p | p | p | p | p | F |
| MD97 | 7/9/2012 | p | p | p | F ³ | p | p | f | f |
| MS12 | 3/7/2012 | p | p | p | F ³ | p | p | n/a | n/a |
| MS99 | 3/7/2012 | p | p | p | F ³ | p | p | n/a | n/a |
| NY06 | 7/11/2012 | p | p | p | p | p | p | p | F |
| NY20 | 8/8/2012 | p | p | F/p ² | F ³ | f | p | p | p |
| NY95 | 7/12/2012 | p | p | p | p | p | p | p | p |
| OH02 | 7/2/2012 | Removed from service | | | | | | | |
| UT97 | 6/12/2012 | p | p | n/a | p | p | p | p | F |
| VT99 | 8/6/2012 | p | p | p | p | f | n/a | n/a | n/a |
| WV99 | 7/5/2012 | F ¹ | p | p | n/a | F | F | n/a | F |

¹ 1135 case temperature reading 175°C and was not repaired

² 2537 flow was 17% low before being corrected

³ Leak checks were high which may be due to incorrect mass flow meter zero offset

2.3 Siting Criteria

Siting criteria is evaluated with regard to obstructions in each of 8 direction (i.e., N, NE, E, SE, S, SW, W, and NW) from the instrument inlet. Inlet heights from the ground are also measured. Results are presented in Table 4. Obstructions are evaluated as pass (p)/fail (f). Deviations from the siting criteria are discussed with the operator during the site survey. Corrective action, when possible, is the responsibility of the site operator and the site supervisor.

Table 4. Siting Criteria Obstructions and Inlet Heights.

| Site | Inlet Height (m) | N | NE | E | SE | S | SW | W | NW |
|------|------------------|---|----|---|----|---|----|---|----|
| CA48 | 3.1 | p | p | p | p | p | p | p | p |
| HI00 | 5.0 | p | p | p | p | p | p | f | p |
| MD08 | n/a | p | p | p | p | p | p | p | p |
| MD96 | 10 | p | p | p | p | p | p | p | p |
| MD97 | 10 | p | p | p | p | p | p | p | p |
| MS12 | 10 | p | p | p | p | p | p | p | p |
| MS99 | 10 | p | p | p | p | p | p | p | p |
| NY06 | 9.1 | p | p | p | p | p | p | p | f |
| NY20 | 4.9 | f | p | p | p | p | f | f | f |
| NY95 | 4.3 | f | p | p | p | p | p | p | f |
| OH02 | 2.5 | p | p | p | p | p | p | p | p |
| UT97 | 8.2 | p | p | p | p | p | p | p | p |
| VT99 | 5.9 | f | p | p | p | p | p | p | p |
| WV99 | 3.2 | p | p | p | p | p | p | p | p |

2.4 Instrument Repairs

In 2012, one instrument (located at VT99) required repairs in order to complete the survey. By comparison, in 2011, seven instruments required repairs in order to complete the survey.

3.0 Training

No formal training sessions were held in 2012. Informal training occurs as part of the site survey with the site liaison observing the maintenance activities, making recommendations. Site surveys at MS99 and VT99 occurred after new operators assumed responsibility for operation of the sites. As such, additional time was spent training the operators.

4.0 Data

AMNet data are evaluated using a series of automated checks, and through manual inspection by the AMNet Site Liaison. Additional information on this process is available in the *Atmospheric Mercury Network Data Management Manual*. Table 5 lists the percent valid data for each site in 2012. Values are presented for each of the three forms of interest, that is, GEM, GOM, and PBM_{2.5}. Two sites, UT97 and WV99, did not meet data quality objectives ($\geq 75\%$ data completeness on annual basis) for 2012.

UT97 experienced two pump failures with the 2537, resulting in significant periods without data. The outage occurred 01-27 January. The second outage occurred 18 August through 17 October.


WV99 experienced severe channel bias between the cartridges from 01 January through 30 May. On 14 October, operations at WV99 were discontinued.

Two other sites experienced periods of downtime greater than 2 weeks, AL19 was shut down from 12 January through 12 March. Heater failure with the 1135 caused NY95 to be out of operation 01-25 August.

Table 5. Percent Valid Data by Site for 2012.

| Site ID | GEM | GOM | PBM |
|---------|-----|-----|-----|
| AL19 | 92 | 92 | 92 |
| FL96 | 88 | 86 | 86 |
| GA40 | 84 | 85 | 85 |
| HI00 | 90 | 86 | 85 |
| MD96 | 98 | 98 | 96 |
| MD97 | 98 | 97 | 97 |
| MS12 | 97 | 94 | 92 |
| MS99 | 96 | 92 | 85 |
| NS01 | 96 | 96 | 96 |
| NY06 | 98 | 94 | 94 |
| NY20 | 94 | 93 | 93 |
| NY95 | 83 | 87 | 80 |
| OK99 | 75 | 89 | 88 |
| UT97 | 86 | 76 | 76 |
| VT99 | 97 | 95 | 96 |
| WI07 | 95 | 95 | 88 |
| WV99 | 38 | 36 | 36 |
| Average | 87 | 88 | 86 |

Appendix – Test Equipment Calibration Documents



Calibration Certificate

| | | | |
|------------------------|-----------------------|-----------------|---|
| Certificate No. | 5006926 | Sold to: | National Atmospheric Deposition Program- NADP - IL |
| Product | Definer 220 High Flow | | 1876 Lewis Road |
| Serial No. | 114711 | | Mt. Hereb, WI 53572 |
| Cal. Date | 7/18/2011 | | USA |

All calibrations are performed in accordance with ISO 17025 at Bios International Corporation, 10 Park Place, Butler, NJ, 07405, 800-663-4977, an ISO 17025:2005 – accredited laboratory through NVLAP. This report shall not be reproduced except in full without the written approval of the laboratory. Results only relate to the items calibrated. This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the Federal Government.

All units tested in accordance with Bios International Corporation test number PR18-13 using high-purity bottled nitrogen or dry filtered laboratory air.

As Received Calibration Data

Technician Sonia Otero
 Lab. Pressure 751 mmHg
 Lab. Temperature 22.5 °C

| Instrument Reading | Lab Standard Reading | Deviation | Allowable Deviation | As Received |
|--------------------|----------------------|-----------|---------------------|------------------|
| 498.96 sccm | 501.995 sccm | -0.6% | 1.00% | In Tolerance |
| 4989.7 sccm | 5007 sccm | -0.35% | 1.00% | In Tolerance |
| 30122 sccm | 30039 sccm | 0.28% | 1.00% | In Tolerance |
| 23.5 °C | 22.5 °C | 1°C | ±0.8°C | Out of Tolerance |
| 750 mmHg | 751 mmHg | -1 mmHg | ±3.5mmHg | In Tolerance |

Bios International Standards Used

| Description | Standard Serial Number | Calibration Date | Calibration Due Date |
|-----------------------|------------------------|------------------|----------------------|
| ML-800-44 | 103521 | 11/10/2010 | 11/10/2011 |
| Precision Thermometer | 305460 | 8/9/2010 | 8/9/2011 |
| Precision Barometer | 431/98-07 | 4/25/2011 | 4/24/2012 |

Bios International • 10 Park Place Butler, NJ 07405 • 800.663.4977 • www.biosint.com



As Shipped Calibration Data

Certificate No. 5021752 Lab. Pressure 760 mmHg
Technician Jacquella Shives Lab. Temperature 22.4 °C

| Instrument Reading | Lab Standard Reading | Deviation | Allowable Deviation | As Shipped |
|--------------------|----------------------|-----------|---------------------|--------------|
| 502.62 sccm | 501.285 sccm | 0.27% | 1.00% | In Tolerance |
| 5009.9 sccm | 5008.95 sccm | 0.02% | 1.00% | In Tolerance |
| 30193 sccm | 30150.5 sccm | 0.14% | 1.00% | In Tolerance |
| 22.4 °C | 22.4 °C | - | ±0.8°C | In Tolerance |
| 760 mmHg | 760 mmHg | - | ±3.5mmHg | In Tolerance |

Bios International Standards Used

| Description | Standard Serial Number | Calibration Date | Calibration Due Date |
|-----------------------|------------------------|------------------|----------------------|
| ML-800-44 | 101897 | 11/16/2012 | 11/16/2013 |
| Precision Thermometer | 305460 | 8/20/2012 | 8/20/2013 |
| Precision Barometer | 2981392 | 6/4/2012 | 6/4/2013 |



Calibration Notes

Bios is an ISO 17025-accredited metrology laboratory. Each Bios primary gas flow standard is dynamically verified by comparing it to one of our laboratory standards, which is a Proven DryCat® Technology volumetric piston prover of much higher accuracy (±0.25 % or less) but of similar operating principles. For this purpose, a flow generator of ±0.10 % or less stability is used. Our laboratory standards are qualified by direct measurement of their dimensions (diameter, length and time) using NIST-traceable precision gauges and instruments, such as depth micrometers and laser micrometers. Calibration Certificates for these gauges and instruments are available upon request. Rigorous analyses of our laboratory standards' uncertainties have been performed, in accordance with The Guide to the Expression of Uncertainty in Measurement (the GUM), assuring their traceable accuracy.
Flow readings in sccm performed at STP of 21.1°C and 760 mmHg.

Technician Notes: _____

David W. Wilson, Chief Metrologist

Bios International, a division of Mesa Laboratories Inc. 10 Park Place Butler, NJ 07405 USA
(973) 492-8400 FAX (973) 492-8270 www.biosint.com www.mesalabs.com Symbol "MLAB" on the NASDAQ



Calibration Certificate

| | | | |
|------------------------|-------------------------|-----------------|---|
| Certificate No. | 5006925 | Sold to: | National Atmospheric Deposition Program- NADP - IL |
| Product | Definer 220 Medium Flow | | 1876 Lewis Road |
| Serial No. | 113878 | | Mt. Hereb, WI 53572 |
| Cal. Date | 7/19/2011 | | USA |

All calibrations are performed in accordance with ISO 17025 at Bios International Corporation, 10 Park Place, Butler, NJ, 07405, 800-663-4977, an ISO 17025:2005 – accredited laboratory through NVLAP. This report shall not be reproduced except in full without the written approval of the laboratory. Results only relate to the items calibrated. This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the Federal Government.

All units tested in accordance with Bios International Corporation test number PR18-13 using high-purity bottled nitrogen or dry filtered laboratory air.

As Received Calibration Data

Technician Sonia Otero
 Lab. Pressure 751 mmHg
 Lab. Temperature 22.6 °C

| Instrument Reading | Lab Standard Reading | Deviation | Allowable Deviation | As Received |
|--------------------|----------------------|-----------|---------------------|--------------|
| 100.32 sccm | 100.415 sccm | -0.09% | 1.00% | In Tolerance |
| 1007.4 sccm | 1009.2 sccm | -0.18% | 1.00% | In Tolerance |
| 5008.3 sccm | 5007.55 sccm | 0.01% | 1.00% | In Tolerance |
| 22.6 °C | 22.6 °C | 0°C | ±0.8°C | In Tolerance |
| 750 mmHg | 752 mmHg | -2 mmHg | ±3.5mmHg | In Tolerance |

Bios International Standards Used

| Description | Standard Serial Number | Calibration Date | Calibration Due Date |
|-----------------------|------------------------|------------------|----------------------|
| ML-800-24 | 100439 | 4/21/2011 | 4/20/2012 |
| Precision Thermometer | 305460 | 8/9/2010 | 8/9/2011 |
| Precision Barometer | 431/98-07 | 4/25/2011 | 4/24/2012 |

Bios International • 10 Park Place Butler, NJ 07405 • 800.663.4977 • www.biosint.com



As Shipped Calibration Data

Certificate No. 5021751 Lab. Pressure 760 mmHg
Technician Jacquella Shives Lab. Temperature 22.4 °C

| Instrument Reading | Lab Standard Reading | Deviation | Allowable Deviation | As Shipped |
|--------------------|----------------------|-----------|---------------------|--------------|
| 100.51 sccm | 100.45 sccm | 0.06% | 1.00% | In Tolerance |
| 1004.9 sccm | 1005.15 sccm | -0.02% | 1.00% | In Tolerance |
| 4998.2 sccm | 5003.65 sccm | -0.11% | 1.00% | In Tolerance |
| 22.4 °C | 22.4 °C | - | ±0.8°C | In Tolerance |
| 760 mmHg | 760 mmHg | - | ±3.5mmHg | In Tolerance |

Bios International Standards Used

| Description | Standard Serial Number | Calibration Date | Calibration Due Date |
|-----------------------|------------------------|------------------|----------------------|
| ML-800-24 | 100439 | 4/24/2012 | 4/24/2013 |
| Precision Thermometer | 305460 | 8/20/2012 | 8/20/2013 |
| Precision Barometer | 2981392 | 6/4/2012 | 6/4/2013 |

Calibration Notes

Bios is an ISO 17025-accredited metrology laboratory. Each Bios primary gas flow standard is dynamically verified by comparing it to one of our laboratory standards, which is a Proven DryCal® Technology volumetric piston prover of much higher accuracy (±0.25 % or less) but of similar operating principles. For this purpose, a flow generator of ±0.10 % or less stability is used. Our laboratory standards are qualified by direct measurement of their dimensions (diameter, length and time) using NIST-traceable precision gauges and instruments, such as depth micrometers and laser micrometers. Calibration Certificates for these gauges and instruments are available upon request. Rigorous analyses of our laboratory standards' uncertainties have been performed, in accordance with The Guide to the Expression of Uncertainty in Measurement (the GUM), assuring their traceable accuracy.

Flow readings in sccm performed at STP of 21.1°C and 760 mmHg.

Technician Notes: _____

David W. Wilson, Chief Metrologist

Bios International, a division of Mesa Laboratories Inc. 10 Park Place Butler, NJ 07405 USA
(973) 492-8400 FAX (973) 492-8270 www.biosint.com www.mesalabs.com Symbol "MLAB" on the NASDAQ

LITHO. IN U.S.A.

CERTIFICATE of CALIBRATION

This is to Certify that

the following described Hamilton Digital Syringe has been calibrated by
Hamilton Company, and is accurate within $\pm 0.5\%$ of full scale reading.

This Digital Syringe, as specified below, has been calibrated as a complete assembly at ambient
pressure. The calibration is performed pursuant to ANSI/NCSL Z540-1-1994, with an unbroken chain
of calibrations traceable to NIST

Capacity 25 μ l

Model 1702RN,25UL

Serial No. 06793 Accuracy 0.086%

Date of Calibration August 19, 2011

Calibrated by Jolene Mayer

HAMILTON
THE MEASURE OF EXCELLENCESM

4970 Energy Way • Reno, Nevada • 89502-4178 • U.S.A.
Telephone +1-775-858-3000 • Fax +1-775-856-7259
Toll Free 800-648-5950

ISO 9001
CERTIFIED

P/N69042 (Rev. G)

NIST test numbers: 821/878169-09 (Mass) S168036 (Temp) 822/785634-10 (Length)

© Giese X411146